



## Introducing EbolaCheck: potential for point-of-need infectious disease diagnosis

Sterghios Athanasios Moschos

**To cite this article:** Sterghios Athanasios Moschos (2015) Introducing EbolaCheck: potential for point-of-need infectious disease diagnosis, Expert Review of Molecular Diagnostics, 15:10, 1237-1240, DOI: [10.1586/14737159.2015.1084228](https://doi.org/10.1586/14737159.2015.1084228)

**To link to this article:** <https://doi.org/10.1586/14737159.2015.1084228>



Published online: 31 Aug 2015.



Submit your article to this journal [↗](#)



Article views: 1000



View related articles [↗](#)



View Crossmark data [↗](#)

# Introducing EbolaCheck: potential for point-of-need infectious disease diagnosis

Expert Rev. Mol. Diagn. 15(10), 1237–1240 (2015)



**Sterghios Athanasios Moschos**

Department of Biomedical Sciences, Westminster Genomic Services, Faculty of Science and Technology, University of Westminster, London, UK  
Tel.: +44 0 20 79 11 50 00  
s.moschos@westminster.ac.uk

The 2013–2015 Ebolavirus disease humanitarian crisis has spurred the development of laboratory-free, point-of-care nucleic acid testing solutions. EbolaCheck is an international consortium of public health, academic and biotechnology industry stakeholders aiming to deliver clinical molecular diagnostic standard-of-care testing suitable for the West African milieu within 12 months. In this article, the current status of the EbolaCheck platform is discussed in the context of the current regulatory framework. Presented here are future goals to achieve differential diagnosis of hemorrhagic fever disease from <5-μl of whole blood samples or mucosal biofluids, in a single tube process, under 40 min and with minimal operator training requirements.

## Background: clinical diagnosis of Ebolavirus disease

Ebolavirus disease (EVD) is an hemorrhagic fever disease (HFD) caused by members of the Filoviridae family of RNA viruses. The filamentous Ebolavirus virion (~90 × 1000 nm) houses a seven gene, ~19 kb genome packed in a nucleoprotein (NP) sheath. Transmission is mediated via the Ebolavirus transmembrane glycoprotein (GP) primarily via macrophage/monocytes. The GP also features immunomodulation, immune evasion and endothelial barrier disruption roles [1]. The monocytic tropism of Ebolavirus mediates proinflammatory responses during replication that amplify infectivity and pathology, collectively resulting in the internal hemorrhage and organ failure characteristic of the later stages of disease [2,3].

Diagnosis is extremely difficult [2,4] as symptoms mimic other HFDs, flu or gastrointestinal infections, which do not preclude Ebolavirus coinfection [4,5]. Transmission risk increases in line with symptom severity, mirroring viremia [6]; presymptomatic patients are not

considered contagious and may remain asymptomatic for up to 21 days [3]. Confirmation of Ebolavirus as the causal disease agent requires clinical molecular diagnostic laboratory solutions. To date, USD 100, <8-hr long, reverse transcription polymerase chain reaction nucleic acid tests (NAT) on RNA extracts from 3.5 ml of whole blood sample (WBS) are the method of choice [7].

However, at the height of the EVD outbreak, the lack of capacity in West Africa required sample shipment overseas, resulting in 3–5-day turnaround times and post-mortem diagnosis [1,8]. The need for a true point-of-need NAT was acute, yet no *in vitro* diagnostic test had received regulatory clearance. ‘Homebrew’ assays were based on Trombley *et al.* (the ‘Trombley’ assays; US Army Medical Research Institute for Infectious Disease; USAMRIID) [9] or Panning *et al.* [10]. These eventually received US FDA emergency use authorization (EUA; EZ1 assay) or were made commercially available under the self-certification CE marking principles (Altona RealStar® Filovirus Screen) [11], respectively.

EXPERT  
REVIEWS

**KEYWORDS:** Ebola • molecular diagnostics • non-profit • nucleic acid testing • point of care • portable

### Molecular diagnostics for infectious diseases at the point of need

Following 9/11 and the subsequent airborne viral disease pandemics, efforts were made to develop decentralized, point-of-care NATs [12]. The resulting solutions, however, were not designed with resource-limited settings in mind [13], despite the ASSURED criteria espoused by the WHO [14]. Thus, the need for a safe, cheap, simple, robust, portable and battery-operated solution remained, presenting an attractive development opportunity for emerging NAT technologies. However, clinical development costs [13], convoluted intellectual property landscapes and industry doubts over outbreak duration and return on investment potential, presented substantial obstacles. Poignantly, despite corporate social responsibility opportunities, to date, all of the major diagnostics manufacturers that engaged in the Ebola response offered primer-probe kits for existing lab-based platforms, or developed 'cassette' kits for existing, closer-to-patient systems. Importantly, these cassette systems maintain for-profit pricing structures for low- and middle-income countries, even following receipt of philanthropic donations in support of their development. The monetization/investment barrier remains cornerstone to both regulator and nongovernment support organization efforts [15].

Yet despite large industry indifference, several academic groups and start up/spinout companies sought to address the point-of-care clinical diagnostic need. However, they faced skepticism from some regulatory bodies regarding manufacturing capacity, quality assurance, commercial launch/support and distribution capability [16]. Thus, little consideration was given to post proof-of-principle, nonprofit production and distribution opportunities similar to regulator-certified, generic pharmaceuticals supply chain models. Under normal circumstances, this would appear appropriate considering the high risk to individual and public health on account of false-positive or false-negative misdiagnosis (WHO category 4 *in vitro* diagnostic classification). However, in August 2014, WHO declared the West African Ebola outbreak as a public health emergency of international concern. Interestingly, this motivated the US FDA to enable EUA approvals; in contrast, the WHO demanded engagement through the full prequalification process. This diverged significantly from the documented successes with other WHO-listed, FIND Diagnostics-vetted, but academic-lead efforts to address neglected disease diagnostics need. The net result was limited performance validation facilitation (access to stored patient samples managed by the WHO) for innovations aiming to address the humanitarian need in the affected countries at the point of care, in lieu for questionable support to preferred lab-based platforms.

### EbolaCheck: the team

The EbolaCheck consortium was formed in response to the August 2014 call of the Research for Health in Humanitarian Crises (R2HC) program, managed by Enhancing Learning and Research for Humanitarian Assistance [17]. The R2HC program aims to improve health outcomes by strengthening the evidence

base for public health interventions in humanitarian crises (visit [18] for more information). The goal of the joint effort between University of Westminster, BioGene Ltd., Public Health England, USAMRIID and the Kwame Nkrumah University of Science and Technology funded through R2HC is to deliver by November 2015, a novel point-of-need NAT solution for simple, rapid and safe patient triage for EVD anywhere in West Africa.

### EbolaCheck: key principles

EbolaCheck can be divided into four sub-systems: the NAT instrument, the EVD assay, the WBS reaction formulation and the reaction consumable. Together, they aim to replace the clinical molecular diagnostic standard of care with a rapid, point-of-need, sample-to-answer format.

### Low cost suitable for West Africa

A simple, patent-protected, energy and engineering-efficient method enables rapid (<2 min), single-tube access to pathogen and host nucleic acids in biofluids with no need for microfluidics. Direct compatibility with standard, cryoprotectable transcription polymerase chain reaction biochemistries further reduces overall cost. Crucially, EbolaCheck will be available to support the on-going, WHO-declared, EVD humanitarian crisis in Africa at cost only.

### Clinical standard of care reliability

The Trombley assay sets for Ebolavirus Zaire GP and NP [9] were migrated to EbolaCheck (Trombley+) to minimize delays, avoid complex licensing negotiations, and on account of emerging field performance evaluation data. Multiplexed use of the Trombley+ assay sets also discriminate vaccinated from infected patients; NP is not found in the two most advanced EVD clinical vaccine candidates [19,20], a problem in on-going vaccination clinical trials pursued by other R2HC-funded programs (Gilbert S, personal communication). USAMRIID have demonstrated performance across five logs of viral RNA genome equivalents (GE) with 100% analytical specificity against 65 other pathogens and analytical sensitivities of 0.001 (NP) and 0.0001 (GP) plaque-forming units per reaction [9]. The roughly 4000 GE/plaque-forming units ratio observed under biosafety level 4 (BSL4) experimentation [21], suggests a lower limit of detection (LLOD) of 10 GE/reaction, or  $10^4$  GE/ml of WBS. Given typical time-to-presentation in autumn 2014 was >3 days post symptom onset, a LLOD goal of  $10^4$  GE/ml WBS was set for the Trombley+ assays on EbolaCheck. Present performance data on surrogate pseudoviral templates indicate nine logs of quantitative linear dynamic range with a lower limit of quantification of 66 GE/reaction and LLOD of six GE/reaction, that is, in line with our performance targets.

### Simple, sample-to-result standard operating procedure

The plethora of reports on 'simple' medical device misuse by end-users in the developed world underscore the importance of ensuring device reliability, particularly with category 4 *in vitro* diagnostic devices operated under significant duress, in

environmentally challenging conditions [13]. The EbolaCheck standard operating procedure consists of:

- reagent unpacking and automated rehydration,
- five-microliter WBS collection by fingertip lancet puncture and MicroSafe<sup>®</sup> capillary collection,
- Sample ejection into the rehydrated consumable,
- lock and loading onto the EbolaCheck instrument, and
- run initiation by touch screen input.

Availability and status of the eight random-access testing stations is visually identified on the front-facing touch screen. Patient status is simply reported as positive, negative or problematic, with the latter indicating a need to repeat the test due to a failure. Full-run kinetics, analytics and diagnostics can be accessed on-screen or over a WiFi connection.

### Safety

The 5 µl WBS requirement of EbolaCheck presents a significant risk reduction to both HCW and HFD patients compared with the closed system, 3.5 ml Vacutainer<sup>®</sup> Eclipse<sup>™</sup> needle and Vacutainer sample standard-of-care protocol. Thermal cycling is expected to destroy EVD [22]; used, sealed consumables are nonetheless discarded as BSL4 clinical waste. The instrument is fully compatible with chlorine dioxide surface sterilization [22] and designed against ingress of liquids or internal condensation [13]. Secure WiFi interface permits remote system checks, maintenance and full reaction data off-boarding. The random access stations also self-diagnose errors and automatically shut down to prevent misdiagnosis.

### Speed

Tests with full personal protective equipment suggest that the EbolaCheck HCW standard operating procedure takes under 2 min to complete by minimally trained individuals, with time to results in <40 min; real-time reaction progression monitoring suggests high viremia-positive results could be called in as little as 20 min.

### Portability

Field experience from in-country Public Health England response teams advised against easily removed, small-form designs, highlighting the need for higher throughput. The ruggedized, 8-well form maintains power supply independence through either mains and/or car battery/alternator power sources. Furthermore, energy consumption modeling indicates solar power supply to be achievable. Design for safety also achieves durability and reliable operation in savannah, coastal and jungle conditions, without corrosion or performance deterioration: simulated environment tests indicate the instrument can complete runs at temperatures as high as 50°C with 98% humidity, and as low as –20°C.

### Development timeline

Prototype design, engineering and assay development were initiated in November 2014. Internal assay standards

containing the Trombley assay targets were developed in MS2 phage icosahedron (Armored RNA<sup>®</sup>) [23] (commercially available) and lipid bilayer-enveloped HIV pseudovirus [24] (open access) formats. Although 26 and 80–100 nm in size, respectively, these represent a vast cadre of viral pathogens. Thus, BSL4 study requirements have been reduced to confirmatory studies using live Ebolavirus, and BSL2 data support EbolaCheck platform utility against other viral pathogens. BSL4 studies are limited to performance evaluation testing against the clinical standard-of-care NAT Trombley assay on culture preparations of Ebolavirus and fresh WBS derived from nonhuman primate models of Ebolavirus infection. In-country testing with fresh or stored patient samples is not expected on account of continued outbreak decline and current WHO priorities to established technologies. However, at least three instruments will be tested in West Africa using mock sample preparations to confirm system operation, portability and reliability in urban, rural and remote environments.

### Future directions

Our early data support multiplexed detection and quantification potential of 3–4 NAT targets in WBS on EbolaCheck. As positive [25] and detrimental [5] coinfections are common among EVD patients, expansion of multiplexing is necessary, but unlikely to exceed concomitant amplification capability need beyond five targets. Field data also indicate mucosal biofluids such as semen [26], ocular fluid [27] and breast milk [28] might be viral depots in convalescence. Interestingly, culturally acceptable alternatives such as saliva [29] and gingival-crevicular fluid [30] might also be of use for HFD diagnosis. Thus, demonstrating EbolaCheck compatibility with these mucosal biofluids will expand the point-of-need monitoring and surveillance capability and introduce the opportunity for needle-free testing. Early feasibility studies indicate that this may enable differential HFD diagnosis with minimal cost of goods increase.

### Concluding remarks

Of the nine EVD NATs that have received FDA EUA to date, three involve complex cartridge/microfluidic systems. Only the 90-min Cepheid Xpert<sup>®</sup> Ebola assay (May 2015) is reasonably priced for the West African milieu at ~USD 20 per test, despite charitable backing. With a comparable assay cadre and LLOD to EbolaCheck, it features a 3 log, nonquantitative dynamic range in highly diluted WBS, requires sample preprocessing, multiple mechanical steps and a separate personal computer and barcode scanner. Despite >10,000 instruments placed worldwide, this WHO-selected platform costs US\$17,000–17,500 to eligible countries. Thus, per-unit scaled production costs are comparable with the current manufacturing cost of EbolaCheck prototypes and the Trombley+ EVD assays. The EbolaCheck consortium has demonstrated that humanitarian crises can motivate efforts to the significant potential benefit of those in need

and leverage development opportunities for appropriately positioned technologies from socially responsible industry with commercial interests in the West. The EbolaCheck consortium is presently seeking charitable support toward scale-up production and delivery of the first differential HFD diagnosis solution, to be provided at cost for any future WHO-declared humanitarian crises.

### Financial & competing interests disclosure

*This work was funded by grant no. 13335 of the Research for Health in Humanitarian Crises (R2HC) Programme managed by ELRHA, funded equally by the Wellcome Trust and DFID. The author has no other relevant affiliations or financial involvement with any organization or entity with a financial interest in or financial conflict with the subject matter or materials discussed in the manuscript apart from those disclosed.*

### References

- Cenciarelli O, Pietropaoli S, Malizia A, et al. Ebola virus disease 2013-2014 outbreak in west Africa: an analysis of the epidemic spread and response. *Inter J Microbiol* 2015;2015:769121
- Kalra S, Kelkar D, Galwankar SC, et al. The emergence of ebola as a global health security threat: from 'lessons learned' to coordinated multilateral containment efforts. *J Global Infect Dis* 2014;6(4):164-77
- Ghazanfar H, Orooj F, Abdullah MA, Ghazanfar A. Ebola, the killer virus. *Infect Dis Pover* 2015;4:15
- Boisen ML, Schieffelin JS, Goba A, et al. Multiple circulating infections can mimic the early stages of viral hemorrhagic fevers and possible human exposure to filoviruses in Sierra Leone prior to the 2014 outbreak. *Viral Immunol* 2015;28(1):19-31
- Gire SK, Goba A, Andersen KG, et al. Genomic surveillance elucidates Ebola virus origin and transmission during the 2014 outbreak. *Science* 2014;345(6202):1369-72
- Towner JS, Rollin PE, Bausch DG, et al. Rapid diagnosis of Ebola hemorrhagic fever by reverse transcription-PCR in an outbreak setting and assessment of patient viral load as a predictor of outcome. *J Virol* 2004;78(8):4330-41
- Spengler JR, McElroy AK, Harmon JR, et al. Relationship between ebola virus real-time quantitative polymerase chain reaction-based threshold cycle value and virus isolation from human plasma. *J Infect Dis* 2015. [Epub ahead of print]
- Tambo E, Ugwu EC, Ngogang JY. Need of surveillance response systems to combat Ebola outbreaks and other emerging infectious diseases in African countries. *Infect Dis Pover* 2014;3:29
- Trombley AR, Wachter L, Garrison J, et al. Comprehensive panel of real-time TaqMan polymerase chain reaction assays for detection and absolute quantification of filoviruses, arenaviruses, and New World hantaviruses. *Am J Trop Med Hyg* 2010;82(5):954-60
- Panning M, Laue T, Olschlager S, et al. Diagnostic reverse-transcription polymerase chain reaction kit for filoviruses based on the strain collections of all European biosafety level 4 laboratories. *J Infect Dis* 2007;196(Suppl 2):S199-204
- Moschos SA. Genomic biomarkers for patient selection and stratification: the cancer paradigm. *Bioanalysis* 2012;4(20):2499-511
- Holland CA, Kiechle FL. Point-of-care molecular diagnostic systems—past, present and future. *Curr Opin Microbiol* 2005;8(5):504-9
- Peeling RW, McNerney R. Emerging technologies in point-of-care molecular diagnostics for resource-limited settings. *Expert Rev Mol Diagn* 2014;14(5):525-34
- Peeling RW, Holmes KK, Mabey D, Ronald A. Rapid tests for sexually transmitted infections (STIs): the way forward. *Sex Transm Infect* 2006;82(Suppl 5):v1-6
- Chua AC, Cunningham J, Moussy F, et al. The case for improved diagnostic tools to control ebola virus disease in West Africa and how to get there. *PLoS Negl Trop Dis* 2015;9(6):e0003734
- Perkins MD, Kessel M. What Ebola tells us about outbreak diagnostic readiness. *Nat Biotechnol* 2015;33(5):464-9
- Enhancing learning and research for humanitarian assistance. Available from: [www.elrha.org/](http://www.elrha.org/)
- Research for health in humanitarian crises. Available from: [www.elrha.org/r2hc/projects](http://www.elrha.org/r2hc/projects)
- Sarwar UN, Costner P, Enama ME, et al. Safety and immunogenicity of DNA vaccines encoding Ebolavirus and Marburgvirus wild-type glycoproteins in a phase I clinical trial. *J Infect Dis* 2015;211(4):549-57
- Regules JA, Beigel JH, Paolino KM, et al. A Recombinant Vesicular Stomatitis Virus Ebola Vaccine - Preliminary Report. *N Engl J Med* 2015; Epub ahead of print
- Rossi CA, Kearney BJ, Olschner SP, et al. Evaluation of ViroCyt(R) Virus Counter for rapid filovirus quantitation. *Viruses* 2015;7(3):857-72
- Cook BW, Cutts TA, Nikiforuk AM, et al. Evaluating environmental persistence and disinfection of the ebola virus makona variant. *Viruses* 2015;7(4):1975-86
- Pasloske BL, Walkerpeach CR, Obermoeller RD, et al. Armored RNA technology for production of ribonuclease-resistant viral RNA controls and standards. *J Clin Microbiol* 1998;36(12):3590-4
- Mather ST, Wright E, Scott SD, Temperton NJ. Lyophilisation of influenza, rabies and Marburg lentiviral pseudotype viruses for the development and distribution of a neutralisation-assay based diagnostic kit. *J Virol Methods* 2014; Epub ahead of print
- Lauck M, Bailey AL, Andersen KG, et al. GB virus C coinfections in west African Ebola patients. *J Virol* 2015;89(4):2425-9
- Christie A, Davies-Wayne GJ, Cordier-Lasalle T, et al. Possible sexual transmission of ebola virus - liberia, 2015. *MMWR* 2015;64(17):479-81
- Varkey JB, Shantha JG, Crozier I, et al. Persistence of Ebola Virus in Ocular Fluid during Convalescence. *N Engl J Med* 2015;372(25):2423-7
- Moreau M, Spencer C, Gozalbes JG, et al. Lactating mothers infected with Ebola virus: EBOV RT-PCR of blood only may be insufficient. *Euro Surveill* 2015;20:3
- Formenty P, Leroy EM, Epelboin A, et al. Detection of Ebola virus in oral fluid specimens during outbreaks of Ebola virus hemorrhagic fever in the Republic of Congo. *Clin Infect Dis* 2006;42(11):1521-6
- Corstjens PL, Abrams WR, Malamud D. Detecting viruses by using salivary diagnostics. *J Am Den Ass* 2012;143(10 Suppl):12S-8S